

CHARGE: 502(a)—the labeling accompanying the article, when shipped, contained false and misleading representations that the article was an adequate and effective treatment for the prevention and correction of jaundice, cirrhosis of the liver, atherosclerosis and diabetes mellitus.

DISPOSITION: 8-27-58. Default—destruction.

5735. Sugar-chek (urine sugar test tape). (F.D.C. No. 41901. S. Nos. 12-980 P, 14-601 P.)

QUANTITY: 385 ctns., each containing 144 retail ctns., at Elgin, Ill.

SHIPPED: 1-25-57 and 2-4-57, from St. Louis, Mo.

LABEL IN PART: "Sugar-Chek * * * (Urine Sugar Test Tape) A specific test for urine glucose (sugar)."

ACCOMPANYING LABELING: (Leaflet in ctn.) "What is a Diabetic."

RESULTS OF INVESTIGATION: Examination showed that the article, while held for sale, was not effective for detecting small amounts of sugar in the urine as claimed in its labeling.

LIBELED: 6-30-58, N. Dist. Ill.

CHARGE: 502(a)—the labeling of the article, while held for sale, contained false and misleading representations that the article was effective for detecting the presence of sugar in the urine and thus acting as a diagnostic sign of diabetes.

DISPOSITION: 7-25-58. Default—destruction.

5736. Slenda-Matic devices. (F.D.C. No. 41794. S. Nos. 25-239 P, 26-551 P.)

QUANTITY: 36 rectangular *Slenda-Matic devices* and 12 contour *Slenda-Matic devices* at Minneapolis, Minn.

SHIPPED: 5-7-58, from Chicago, Ill., by Grayline Co.

LABEL IN PART: (Ctn.) "Rectangular Holco Slenda-Matic Distributed By S. S. Hollender, Inc., Chicago, Illinois" and "Contour Holco Slenda-Matic"; (metal plate on each device) "Slenda-Matic Manufactured by Grayline Company, Chicago 33, Illinois * * * Serial No."

ACCOMPANYING LABELING: Leaflet entitled "Holco Slenda-Matic Vibrating Machine."

RESULTS OF INVESTIGATION: The labeling indicated that each device was a wooden frame box upholstered with polyurethane foam and covered with vinyl material, and that a rheostat-control electric motor providing vibration was enclosed in each box.

LIBELED: 6-4-58, Dist. Minn.

CHARGE: 502(a)—the labeling of the devices, when shipped, contained false and misleading representations that the devices were capable of providing an adequate and effective treatment for improving circulation, easing tensions, reducing weight, regaining a firm, trim, slender figure and removing fatty tissues.

DISPOSITION: 7-22-58. Consent—claimed by E. B. Meyrowitz, Inc., St. Paul, Minn., and relabeled.

5737. Ozone generator. (F.D.C. No. 41454. S. No. 41-302 P.)

QUANTITY: 11 devices at Seattle, Wash.

SHIPPED: 10-21-57, from Germany.